



Ref: INS/F-5791/2017-2018(Y)

Date : 06/12/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Hybridization oven (Specifications Attached)	any	3.00	Nos.

Remarks:The Tenders can also to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".

2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**

3. Country of origion

4. Place of Shipment

5. Name, address,telephone number and fax numbers of your bankers.

6.Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).

7.**Please Indicate frieght and insurance charges seperately.**

8.Please Indicate the Warranty Period and Delivery Period.

9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **12/12/2017 till 5.30 p.m.**

Important Note

10.Shipping by AIRFREIGHT ONLY-CARGO MODE

11.Port of Entry for Customs clearance is -BANGALORE

12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.

Institute For Stem Cell Biology and Regenerative Medicine
Autonomous Institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5791/2017-2018(Y)

Date : 06/12/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to be 'J. S. R.' with a flourish underneath.

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



Ref No: INS/F-5791//2017-2018(Y)

OPTIN - I

SPECIFICATIONS FOR HYBRIDIZATION OVEN

- 3.5" Large display touch screen & graphical control interface
- User temperature calibration
- Broad temperature control range
- Timer with alarm function
- Safety device in the event of system malfunction
- Safety door switch device
- Full accessories available

Cat. No.	MO-A01
Display	3.5" 64K color TFT display
Controller	32-bits Microprocessor
Control interface	Touch screen & graphical interface
Timer	9999 (min) with alarm / Continuous
Temperature control range	Ambient +5°C to 85°C
Temperature increment	0.1°C
Temperature uniformity	± 0.2°C at 37°C
Temperature accuracy	± 0.2°C at 37°C
User temperature calibration	Yes
Inner chamber dimensions (WxLxH)	13.39"x8.86"x10.24" (340x225x260mm)
Inner chamber volume	Approx. 5.28gal (20L)
Dimension (WxLxH)	17.4"x18.19"x17.8" (442x462x452mm)
Platform (WxL)	N/A
Data logging	RS-232
Weight	Approx. 52.9lb (24kg)
Rated voltage	110 / 220V~ 50 / 60Hz (Selectable)
Construction	Painted iron metal with ABS front door
Safety device	Safety door switch
	Thermal safety switch
	Auto shut off upon fan failure
	Power failure auto recovery
Certifications	FCC & CE



SPECIFICATIONS FOR HYBRIDIZATION SHAKING OVEN

- 3.5" Large display touch screen & graphical control interface
- User temperature calibration
- Broad temperature control range
- Timer with alarm function
- Safety device in the event of system malfunction
- Safety door switch device
- Full accessories available

Cat. No.	MO-AOR
Display	3.5" 64K TFT display
Controller	32-bits Microprocessor
Control interface	Touch screen & graphical interface
Rotisserie function	Yes
Rotisserie speed / Inc.	5-100rpm / 1rpm
Timer	9999 (min) with alarm / Continuous
Shaker motion	Orbital clockwise / counterclockwise
Shaker speed	0-200rpm
Temperature control range	Ambient +5°C to 85°C
Temperature increment	0.1°C
Temperature uniformity	± 0.2°C at 37°C
Temperature accuracy	± 0.2°C at 37°C
User temperature calibration	Yes
Inner chamber dimensions (WxLxH)	13.39"x8.86"x10.24" (340x225x260mm)
Inner chamber volume	Approx. 5.28gal:1 (20L)
Dimension (WxLxH)	17.4"x18.19"x17.8" (442x462x452mm)
Platform (WxL)	10.63"x7.87" (270x200mm)
Data logging	RS-232
Weight	Approx. 63.9lb (29kg)
Rated voltage	110V / 220V~ 50 / 60Hz (selectable)
Construction	Painted iron metal with ABS front door
Safety device	Safety door switch
	Thermal safety switch
	Auto shut off upon fan failure
	Power failure auto recovery



Certifications	FCC & CE
----------------	----------

ACCESSORIES	
MO-01-PF	1 layer of stainless steel mesh plate with 4 plate holders
MO-HY-8RT	40mm tube rotisserie for 8 tubes
MO-HY-16RT	50ml conical tube rotisserie for 16 tubes
MO-HY-24RT	15ml conical tube rotisserie for 24 tubes
MO-SEESAW	Rocking shaker platform for MO-ARK
MO-P2320	Addition 23x20cm platform for orbital, and reciprocal shaker (with 4 of attachable 10cm pillars)
SI-200-08	Flask Holder, 50ml
SI-200-09	Flask Holder, 125ml
SI-200-10	Flask Holder, 250ml
SI-200-11	Flask Holder, 500ml
MO-BT40X150	Glass tube 40x150mm(diameter x length)
MO-BT40X200	Glass tube 40x200mm(diameter x length)
MO-BT40X300	Glass tube 40x300mm(diameter x length)

