

Ref: INS/F-6089/2017-2018(N)

Date : 08/01/2018

### ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Probe Sonicator (Specificalitons Attached)	any	1.00	Nos.

**Remarks:The Tenders to be quoted in foreign currencies & any other currencies approved/traded by RBI - USD/Euro/JPY/GBP/SGD/CAD/INR.**

Please also indicate specifically the following :\*\*inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification. \*\*

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".

2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**

3. Country of origion

4. Place of Shipment

5. Name, address,telephone number and fax numbers of your bankers.

Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).

7.**Please Indicate frieght and insurance charges seperately.**

8.Please Indicate the Warranty Period and Delivery Period.

Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **22/01/2018 till 5.30 p.m.**


Important Note

10.Shipping by AIRFREIGHT ONLY-CARGO MODE

11.Port of Entry for Customs clearance is -BANGALORE

2.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.





Ref: INS/L-6089/2017-2018(N)

Date : 08/01/2018

**12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.**

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell  
Biology and Regenerative Medicine

  
08/01/18  
Namratha A

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in





**Ref No: INS/F-6089/2017-2018(X)**

### **Specifications of Probe Sonicator**

1. The system should have microprocessor based Controller.
2. The system should have automatic tuning and frequency control
3. The system should have option to set the time and amplitude
4. The system should be capable of simultaneous monitoring and controlling of both the temperature of the sample and the amount of energy that is being delivered
5. The system should have pulse mode with full function ON and OFF cycle. Continuous mode up to 30 minutes or preferably more.
6. The system should have sound proof enclosure and jack stand
7. Frequency : Up to 40 kHz or more
8. Processing Capacity: up to 1000 ml
9. Digital amplitude / Intensity control : Output amplitude can be set from 10 to 100%
10. Display: Energy monitor, Watt meter, Amplitude, Time indicator, Pulser and Temperature.
11. Probes: suitable for Processing Volume 150  $\mu$ l to 5 ml and 10 ml to 250 ml
12. The system should be quoted with multiple sample processing probes with all kind of tips.
13. The system should have warranty of 3 years.
14. The system should be CE/ISO certified.
15. The system should operate on 230V,50 Hz
16. The system should be supplied with all the accessories required to function.

