



Ref: INS/F-6215/2017-2018(Y)

Date : 12/02/2018

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

| Sl.No | Cat.No | Item Description | Make/Model | Item Qty | UOM |
|-------|--------|---|------------|----------|-----|
| 1 | | Air Compressor Oil free Scroll Type (As per specification attached) | Any | 1.00 | No. |

Remarks: The Tenders can also be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

NOTE:inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification.****

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
3. Country of origion
4. Place of Shipment
5. Name, address,telephone number and fax numbers of your bankers.
- 6.Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
- 7.**Please Indicate frieght and insurance charges seperately.**
- 8.Please Indicate the Warranty Period and Delivery Period.
- 9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **05/03/2018 till 5.30pm**

Important Note

- 10.Shipping by AIRFREIGHT ONLY-CARGO MODE
- 11.Port of Entry for Customs clearance is -BANGALORE
- 12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.





Ref: INS/F-6215/2017-2018(Y)

Date : 12/02/2018

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advalorem duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in blue ink, appearing to read 'Yesu R'.

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in



INSTITUTE FOR STEM CELL BIOLOGY AND REGENERATIVE MEDICINE

Ref No: INS/F-6215/2017-2018(Y)

Specification for Oil free Scroll Air Compressor

- System should be oil free, scroll type and suitable for Electron Microscopy instruments.
- System should have working FAD (LPM): $250 \pm 10\%$ and working pressure 8 Bar.
- System should have working motor type 2.2kWatt, totally enclosed 4 pole
- System should have starting system: - Direct online with magnetic switch
- System should contain Dryer.
- System should have line filter G 24X EA 1Micron.
- System should have line filter G24 Y IA 0.01 Micron.
- Inbuilt air receiver tank capacity should be 20 liters.
- System should be supplied with 100 liter capacity, stainless steel receiver tank.
- System should have electronic Condensate drain unit for 100 liter receiver tank
- System should have ground vibration less than 55db A.
- System should have F class insulation.
- System should have Anti-overheat protection.
- System should have Anti-overcurrent protection.
- System should include all Accessories, along with the flexible hose pies to interconnect compressor, drier and receiver tank, with appropriate connectors.
- System should have ISO/CE Certified.
- System should have 3 year warranty.
- The operating Voltage should be 3 Phase (415 Volt,50 Hz).

