



Ref: INS/L-5506/2017-2018(Y)

Date : 12/10/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Magnetic stirrer hot plate (As per the specification attached)	Any	4.00	µg
2		Tube Roller (As per the specification attached)	Any	2.00	Nos.
3		Heating Block (As per the specification attached)	Any	2.00	Nos.

Remarks :

Note :Quotation in USD, GBP, YEN, EURO, CND, SGD, ect will be acceptable.

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for _____" Ref No : _____. The bids should be addressed and to be mailed to "**THE HEAD-PURCHASE**". The bids are liable to be rejected if the sealed envelope is not addressed to "**THE HEAD-PURCHASE**" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS
26/10/2017 till 5.30 p.m.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5.All duties,taxes,surcharge and cess as currently applicable must be sated in your quotation,seperately.Otherwise your quote is liable to be rejected.

6.Your quotation should indicate delivery period & Warranty period.

7.Delivery to be made to our Stores.Please indicate charges,if any extra.Transit Insurance should be done upto inStem Stores.

8.If you are unable to supply the quality,specifications or brand as mentioned in our enquiry,Please state so and then offer alternative to quality/Specifications.

9.Payment :within one month after delivery & acceptance/satisfactory installation.

10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11.Since we are a public funded research institution,we are exempted from paying Customs Duty(Except advolerum duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments,technical instruments,equipments(including computers),their accessories,spares,consumables and software.Hence,please offer your prices taking this option into consideration.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5506/2017-2018(Y)

Date : 12/10/2017

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in





inStem

Institute For Stem Cell Biology And Regenerative Medicine

Ref No: INS/L-5506/2017-2018(Y)

Magnetic Stirrer hot plate

The system should be Surface type and should be ceramic
The system should have the separate switch/control for stirrer and heater

The system should have the speed range between 60 – 1200 RPM

The system should be able to reach a maximum temperature 300 c

The system should be able to accommodate a maximum 2/3 liters' flask

The system should be able to hold the maximum load 10kg

Power 230v/50hz

The system should be CE/ISO Certified

The system should have the warranty for 3 years

The system should be supplied with all the accessories required to function

Heating Block

The system should have the temperature range ambient 5 degrees to 150 degrees

The system should have the LED/LCD Display

The system should have Digital microprocessor based pid controller

The system should have the maximum load capacity of 30 Lbs

The system top material should be made up of ceramic

The system should have the accuracy of 0.1°C

The system should have the temperature calibration

The system should be supplied with suitable blocks 1.5/2 ml – 20 wells & 0.5 ml

-

30 wells

The system should be CE/ISO Certified

The system should have the warranty for 3 years

Power 230v/50hz

The system should be supplied with all the accessories required to function.

Tube Roller

The system should have the Orbital motion type

The system should have the speed range of 5 – 60 RPM

The system should have the tilt angle of 0 to 90 °C

The system should have the microprocessor based PID controller

The system should have digital speed control even when it is rotating in low speed

The system should have the self-diagnostic function to identify the errors

The system should have the timer range 1 min to 99 hrs

The system should have the rubber feet to have the minimum vibration and noise



The system should be suitable to hold 1.5ml, 5ml, 15ml and 50ml tubes
The system should be able to rotate maximum 100 test tubes

The system should be supplied with optional dual holder accessory.

The system should be CE/ISO Certified

The system should have the warranty for 3 years

Power 230v/50hz

The system should be supplied with all the accessories required to function.

