

Ref: INS/F-5751/2017-2018(Y)

Date : 06/11/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Co2 incubator (Specifications Attached)	any	7.00	Nos.

Remarks:The Tenders to be quoted in foreifn currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
 2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
 3. Country of origion
 4. Place of Shipment
 5. Name, address,telephone number and fax numbers of your bankers.
 6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
 - 7.**Please Indicate frieght and insurance charges seperately.**
 8. Please Indicate the Warranty Period and Delivery Period.
 9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **28/11/2017 till 5.30 p.m.**
- Important Note
10. Shipping by AIRFREIGHT ONLY-CARGO MODE
 11. Port of Entry for Customs clearance is -BANGALORE
 12. Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5751/2017-2018(Y)

Date : 06/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advalorem duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read "Yesu R".

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in





Ref No: INS/F-5751/2017-2018(Y)

Specifications for Air jacketed CO₂ incubator

1. The system should be Air jacketed CO₂ incubator.
2. The system should be microprocessor controlled CO₂ incubator It should have touch screen LCD controller with user-friendly control system and GUI.
3. The system should have the capacity of 170 Lt ($\pm 5\%$)
4. The system should have two doors and inner door should be glass door to seal the chamber with Silicon gasket.
5. The system should have the seamless stainless steel (SS) inner chamber.
6. The system should be supplied with minimum 3 nos of adjustable SS shelves.
7. The system should have the Infrared CO₂ sensor with the CO₂ range 1-20% with accuracy 0.1%.
8. The system should have the Operating temperature range from ambient to 50°C. The fluctuation/variation at 37°C should not be more than 0.1°C.
9. The temperature accuracy and uniformity should be less than 0.1 °C
10. The system should have the tracking alarm $\pm 1^\circ\text{C}$ deviation from set point.
11. Sensor must be able to withstand high temperatures during direct heat sterilization
12. The system should have Built in on-demand high temperature decontamination cycle (140°C and above)
13. The system should have the option to mute the audible alarms
14. The system should be suitable for single phase 230 V, 50 Hz Power supply.
15. The system should have onboard Diagnostic interface to show system parameters and functions.
16. The system should be equipped with buzzer and visual alarm for power failure, CO₂, and temperature deviation against the setup temperature.
17. The Co₂ incubator should be quoted with a suitable stackable kit with the main offer.
18. The system should operate at 230V, 50 Hz AC power.
19. The system should be supplied with all the accessories required to function.
20. The system should be offered with minimum 3-year warranty.
21. The System should be CE/ISO Approved.
22. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).

