



Ref: INS/F-5754/2017-2018(Y)

Date : 10/11/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Table top Refrigerated Centrifuge (As per specification attached)	Any	3.00	Nos.

Remarks:The Tenders to be quoted in foreifn currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".

2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**

3. Country of origion

4. Place of Shipment

5. Name, address,telephone number and fax numbers of your bankers.

6.Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).

7.**Please Indicate frieght and insurance charges seperately.**

8.Please Indicate the Warranty Period and Delivery Period.

9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **27/11/2017 till 5.30 p.m.**

Important Note

10.Shipping by AIRFREIGHT ONLY-CARGO MODE

11.Port of Entry for Customs clearance is -BANGALORE

12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous Institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5754/2017-2018(Y)

Date : 10/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'YESU R'.

YESU R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in





INS/F-5754/2017-2018(Y)

Specification of Table top Refrigerated Centrifuge

1. The system should have a maximum speed of 14000 RPM.
2. The system should have LCD Display with Microprocessor based control System.
3. The system should have Direct, Brushless induction, low profile motor drive system
4. The system should have Imbalance Detection system.
5. The system should have Soft-touch one-finger lid closure for easy operation
6. The system should have temperature control ranging from -10° C to 40° C and quick precooling
7. The system should have Pre-cooling system with Direct Button.
8. The system should have Timer Range from 35 secs to 9:50 hrs, and continuous run function.
9. The system should have Automatic lid opening at the end of the run to prevent the sample and allow easy access to samples
10. The system should have Power failure and Temperature Alarm
11. The system should have the Built-in condensation chamber to avoid water accumulation in the rotor chamber
12. The system should have soft brake to protect sensitive samples.
13. The system should have option for auto power shut off after 8 hr of non-use to extend the compressor life.
14. The system should be supplied with 24x1.5 and 2.0ml and 15ml/50ml Falcon tubes and plate rotors. Fully autoclavable and aerosol tight metallic rotor with metallic lid for safe centrifugation of hazardous samples.
15. The system should be power efficient and low power consumption.
16. The system should be supplied with all the accessories required to function. Compliance to each of the above points should be separately indicated and evidence presented for each of them (Product brochures should be highlighted wherever required).
17. The supplier shall be responsible to pay the amount for any damage to the sample due to breakdown during warranty period.
18. The system should be supplied with 3 years warranty.
19. The system should be CE/ISO Certified
20. System should be suitable for high capacity rotors upto 750ml and swinging bucket rotors/plate rotors
21. The offer should include the cost of rotors as optional items

