

Ref: INS/F-5756/2017-2018(Y)

Date : 07/11/2017

### ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Magnetic Stirrer with Hot Plate (As per specification attached)	Any	8.00	Nos.

**Remarks:The Tenders to be quoted in foreifn currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.**

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
  2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
  3. Country of origion
  4. Place of Shipment
  5. Name, address,telephone number and fax numbers of your bankers.
  6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
  - 7.**Please Indicate frieght and insurance charges seperately.**
  - 8.Please Indicate the Warranty Period and Delivery Period.
  - 9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **23/11/2017 till 5.30 p.m.**
- Important Note
- 10.Shipping by AIRFREIGHT ONLY-CARGO MODE
  - 11.Port of Entry for Customs clearance is -BANGALORE
  - 12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



**Institute For Stem Cell Biology and Regenerative Medicine**  
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5756/2017-2018(Y)

Date : 07/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell  
Biology and Regenerative Medicine

  
Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in



Institute for Stem Cell Biology and Regenerative Medicine  
Autonomous institute of the Dept of Biotechnology, Govt. of India.



INS/F-5756/2017-2018(Y)

### **Magnetic Stirrer with Hot plate**

1. The system should have both stirrer and heater and should have separate control.
2. The should have hotplate and should be ceramic coated.
3. The system should have the speed range of 60 – 1200 RPM with accuracy of  $\pm 2$  rpm.
4. The system should have LCD/LED display for indication of Temperature and RPM.
5. The system should have temperature sensor to measure temperature.
6. The system should have the Maximum temperature 300°C with minimum 750W heater.
7. The system should have microcontroller based temperature control Option for temperature accuracy of  $\pm 1^\circ\text{C}$ .
8. The system should be able to accommodate maximum 2/3 liters' flask and should have stirring capacity of 10Lt water.
9. The system should have an independent safety circuit with separate circuit switches to cut off the heating if the normal hotplate temperature is exceeded by 20°C
10. The system should be suitable for Single phase 230v/50hz Power supply.
11. The system should be CE/ISO Certified
12. The system should have minimum 3-year warranty.
13. The system should be supplied with all the accessories required to function

