



Ref: INS/F-5761/2017-2018(Y)

Date : 06/11/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Liquid nitrogen Dewar 120 liter as per specifications attached.	any	1.00	Nos.

Remarks: The Tenders to be quoted in foreifn currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".

2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**

3. Country of origion

4. Place of Shipment

5. Name, address,telephone number and fax numbers of your bankers.

6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).

7. **Please Indicate freight and insurance charges seperately.**

8. Please Indicate the Warranty Period and Delivery Period.

9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **22/11/2017 till 5.30 p.m.**

Important Note

10. Shipping by AIRFREIGHT ONLY-CARGO MODE

11. Port of Entry for Customs clearance is -BANGALORE

12. Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5761/2017-2018(Y)

Date : 06/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'Yesu-R'.

Yesu-R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



Institute for Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Dept of Biotechnology, Govt. of India.



INS/F-5761/2017-2018(Y)

Specification for Liquid Nitrogen Dewar 120 liters

- 1) System should have all welded stainless steel construction, triple wall vacuum insulated system, which is compatible for liquid Nitrogen storage.
- 2) System should have capacity of 120 liters $\pm 5\%$.
- 3) System should be equipped with heavy duty castors up to 170 liters with Tow bar handles.
- 4) System should Twin liquid fill/decant valves
- 5) System should have an integral pressure build vaporizer with automatic pressure regulation
- 6) System should be equipped with an electronic level gauge which gives an accurate reading of Dewar contents.
- 7) System should equipped with a safety dispensing head incorporating fill, vent and dispense valves, safety reliefs valves and pressure gauges
- 8) System should have liquid nitrogen evaporation rate less than 0.8% per day.
- 9) System should have Maximum allowable working pressure of 3BarG
- 10) System should have 1/2" BSPPF outlet.
- 11) System should include all Accessories, along with the flexible hose pipes with appropriate connectors.
- 12) System should have Color-coded, labelled valves for easy identification.
- 13) System should have 5-year vacuum warranty and for other parts 3-year Warranty.
- 14) System should have complied with Pressure Equipment Directive or equivalent.
- 15) System should be ISO/CE certified.
- 16) Installation of Liquid nitrogen Dewar should be a part of the offer.

