



Ref: INS/F-5763/2017-2018(Y)

Date : 10/11/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Electrophoresis system with horizontal gel station (Specifications Attached)	any	10.00	Nos.

Remarks:The Tenders to be quoted in foreifn currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
 2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
 3. Country of origion
 4. Place of Shipment
 5. Name, address,telephone number and fax numbers of your bankers.
 6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
 - 7.**Please Indicate frieght and insurance charges seperately.**
 8. Please Indicate the Warranty Period and Delivery Period.
 9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **27/11/2017 till 5.30 p.m.**
- Important Note
10. Shipping by AIRFREIGHT ONLY-CARGO MODE
 11. Port of Entry for Customs clearance is -BANGALORE
 12. Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5763/2017-2018(Y)

Date : 10/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advalorem duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Institute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'Yesu R'.

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in





INS/F-5763/2017-2018(Y)

Specifications for Electrophoresis Unit

- Should have control of constant voltage output as well as constant current output. The Output Voltage should be in the range 10 –300 V, adjustable with step of 1 V. The output current from system should be in range of 4-400 mA, adjustable with step of 1 mA.
- The system should have timer range from 1 min to 99 hr 59 min.
- The power pack system unit should have 4 paired banana jacks at the output terminal
- The system should have Manual programming of voltage, current and time parameter. The parameter should be changeable in the power pack without interrupting the run.
- The system should have inbuilt memory to store the frequently used programs
- The system should have safety features like No load, resistance change, overvoltage, over-temperature detection and overload and short circuit and Input power supply protection.
- The system should be suitable for 230 V AC, 50Hz power supply.
- The system should be supplied with the suitable accessories required to function.
- The system should be CE / ISO certified.
- The system should have minimum 3 years warranty.

For electrophoresis Tank-

- The system should have Gel tank and its lid with the power cables
- The system should have a buffer volume of 750 ml to run 2 gels and 950 ml to run 4 gels.
- The system should have a mini Gel format size compatibility.
- The system should be compatible to run either precast or hand-cast gels.
- The system should be supplied with the buffer dam of at-least 2 quantity.
- The system should be supplied with the suitable accessories required to function.
- The system should be CE / ISO certified.
- The system should have minimum 3 years warranty.

