



Ref: INS/F-5769/2017-2018(Y)

Date : 30/11/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		UV-Vis spectrophotometer (Specifications Attached)	any	1.00	Nos.

Remarks:The Tenders to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
3. Country of origion
4. Place of Shipment
5. Name, address,telephone number and fax numbers of your bankers.
6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
- 7.**Please Indicate frieght and insurance charges seperately.**
- 8.Please Indicate the Warranty Period and Delivery Period.
- 9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **11/12/2017 till 5.30 p.m.**

Important Note

- 10.Shipping by **AIRFREIGHT ONLY-CARGO MODE**
- 11.Port of Entry for Customs clearance is **-BANGALORE**
- 12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5769/2017-2018(Y)

Date : 30/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advalorem duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Institute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'Yesu R'.

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in



Ref No: INS/E-5769/2017-2018(Y)

Specifications of NANO DROP

The system should have the UV VIS spectral range of 190 -840 nm

The light source of the system should be Xenon flash lamp should have very long life

The system should option for both pedestal, cuvette measurement.

The system should be supplied with the necessary PC with the instrument

The system should have the Pre-programmed methods to measure absorbance for RNA,Proteins for quick start.

It should have the Fast measurement time of less than 10 seconds

The system should accept 10mm cuvette.

The system should have option for measuring kinetics with adjustable run/wait time.

The system should have option for Blank,Re-Blank.

The system should have the wavelength accuracy of ± 1 nm

The system should have absorbance accuracy of ± 5 nm.

The system have absorbance range of 0.5 to 300.

The system should be able to measure the minimum sample volume.

The system should be supplied with pedestal cleaning solutions.

The system should be supplied with calibration solutions.

The system should support Data Exporting option (Excel,notepad and PDF Formats, etc)

The system should be supplied with all the accessories required to function

The system should have the warranty for 3 years

The system should be included with User-friendly software

The system should be CE/ISO Certified

