



Ref: INS/F-5774/2017-2018(Y)

Date : 30/11/2017

**ENQUIRY**

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Digital analytical balance (Specifications Attached)	any	4.00	Nos.

**Remarks:The Tenders to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.**

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
3. Country of origion
4. Place of Shipment
5. Name, address,telephone number and fax numbers of your bankers.
6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
- 7.**Please Indicate frieght and insurance charges seperately.**
- 8.Please Indicate the Warranty Period and Delivery Period.
- 9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **11/12/2017 till 5.30 p.m.**

Important Note

- 10.Shipping by AIRFREIGHT ONLY-CARGO MODE
- 11.Port of Entry for Customs clearance is -BANGALORE
- 12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



**Institute For Stem Cell Biology and Regenerative Medicine**  
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-5774/2017-2018(Y)

Date : 30/11/2017

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell  
Biology and Regenerative Medicine

  
Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



**Ref No: INS/F-5774/2017-2018(Y)**

### **Analytical balance**

1. Capacity : 150g
2. Minimum display 0.0001g
3. Repeatability 0.0001g
4. Tare function should be available
5. Hold function to lock the weight
6. It should have round weighing platform, 3 door draft shield glass chamber providing 8 inches height from pan to top of glass door
7. MEMORY function to keep the last weight in memory
8. The scale should have inbuilt rechargeable battery backup for minimum of 8 hrs
9. Internal Calibration option should be available
10. Response Time : Approx 3 Seconds
11. Operating Temperature : 4 to 40 deg C
12. Power requirement 230V, 50 Hz
13. Display: A liquid crystal indicator with the attractive light
14. Power saving feature: the backlight automatically shuts off when idle and lights up again when weighing operations resume
15. Low Battery indication should be available while operating on battery
16. The reading should get locked automatically at stable weight and there should be an indication of Stable weight
17. The display should shut off when idle and lights up again when weighing operations resume
18. Data interface RS 232 bidirectional

### **Precision balance**

1. Capacity : 450g
2. Minimum display 0.0001g
3. Repeatability 0.0001g
4. Linearity 0.002g
5. Tare function should be available
6. Hold: function to lock the weight.
7. MEMORY function should be available to keep the last weight in memory.
8. The scale should have inbuilt rechargeable battery backup for minimum of 8 hrs.
9. Internal Calibration feature should be available
10. Response Time : Approx 5 Seconds
11. Operating Temperature : 4 to 40 deg C
12. Power saving feature: the backlight automatically shuts off when idle and lights up again when weighing operations resume
13. Display: A liquid crystal indicator with the attractive light
14. Power 230v/50hz
15. Low Battery indication should be available while operating on battery
16. The reading should get locked automatically at stable weight and there should be an indication of Stable weight
17. The display should shut off when idle and lights up again when weighing operations resume.
18. Data interface RS 232 bidirectional

