

Ref: INS/F-5971/2017-2018(Y)

Date : 03/01/2018

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Chemiluminiscent Imaging System (Specifications Attached)	any	1.00	No.

Remarks:The Tenders to be quoted in foreign currencies & any other currencies approved/traded by RBI - USD/Euro/JPY/GBP/SGD/CAD/INR.

Please also indicate specifically the following :**inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification. **

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".

2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**

3. Country of origion

4. Place of Shipment

5. Name, address,telephone number and fax numbers of your bankers.

6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).

7.**Please Indicate frieght and insurance charges seperately.**

8.Please Indicate the Warranty Period and Delivery Period.

9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **16/01/2018 till 5.30 p.m.**

Important Note

10. Shipping by AIRFREIGHT ONLY-CARGO MODE

11. Port of Entry for Customs clearance is -BANGALORE

12. Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: **INS/F-5971/2017-2018(Y)**

Date : 03/01/2018

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

.inStem is exempted in paying customs duty(except advalorem duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Institute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'Yesu R'.

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in





Specifications for the Chemiluminescent Imaging System:

- ☐ Flexible imaging options—capture visible and near-infrared fluorescence-based western blots and gels, chemiluminescent western blots, and colorimetric stained protein and DNA Gels
- ☐ It should be designed to deliver high quality images every time with auto-alignment, auto exposure, auto-zoom, and auto-focus
- ☐ Mechanical sample alignment
- ☐ The system should image and analyse chemiluminescent western blots and stained Protein (Coomassie, silver, sypro, etc), DNA (EtBr, Sybr, etc) gels and fluorescence from visible to NIR range.
- ☐ The system should have five channel fluorescence and four channel multiplexing.
- ☐ The system should have a Smart Exposure Technology which rapidly determines optimal exposure time. It should help in minimising the potential for over- or underexposed images, and the need to repeat exposures to get the desired signal
- ☐ The system should have a sensitive >8 megapixel (MP) cooled CCD camera
- ☐ The system should have a Large Sample Area and should accommodate 4 or more gels/Blots at the same time. The dimensions of the imaging area should be (W) 230mm x (D) 184mm.
- ☐ The system should do Auto Zoom and Auto Focus.
- ☐ The system should utilize a transilluminator based on green LEDs.
- ☐ The system should have onboard Auto Analysis software for instantaneous lane and band identification and molecular weight marker overlay. Quantitation and densitometry analysis should also be performed directly on the instruments
- ☐ The system should have a cloud connectivity.
- ☐ Display by 12.1 touch screen
- ☐ Warranty 3 years

