

Ref: INS/F-6085/2017-2018(N)

Date : 02/01/2018

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Fast Protein Liquid Chromatography (Specifications Attached)	any	1.00	Nos.

Remarks:The Tenders to be quoted in foreign currencies & any other currencies approved/traded by RBI - USD/Euro/JPY/GBP/SGD/CAD/INR.

Please also indicate specifically the following :**inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification. **

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".

2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**

3. Country of origion

4. Place of Shipment

5. Name, address,telephone number and fax numbers of your bankers.

Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).

7.**Please Indicate frieght and insurance charges seperately.**

8.Please Indicate the Warranty Period and Delivery Period.

Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by
19/01/2018 till 5.30 p.m

Important Note

10.Shipping by AIRFREIGHT ONLY-CARGO MODE

11.Port of Entry for Customs clearance is -BANGALORE

2.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-6085/2017-2018(N)

Date : 02/01/2018

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advalorem duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

Handwritten signature of Namratha A. in black ink, dated 02/01/18.

Namratha A



GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



Ref No: INS/F-6085/2017-2018(Y)

Specifications for Fast protein liquid chromatography (FPLC)

1. A computer operated FPLC machine for various chromatographic procedures involved in protein purification.
2. System should support the application like gel filtration, Ion exchange chromatography and hydrophobic interaction chromatography.
3. Reliable piston pumps for gradients. Flow rates from 0 to 10 ml/ or greater than 10ml
4. The system should have maximal Pressures of up to or greater than 10mPa
5. Operational at room temperature and cold room
6. Reliable inline UV Detector with multi wavelength (up to 3 wavelength) preferably LED based.
7. System should have UV range from 200 to 800nm.
8. Inline pH with the range from 0 to 20 and conductivity monitor
9. System should be supplied with at least 3 sample loop
10. Valve for Multiple column option
11. Sample loading pump and valve
12. Flexible fraction collector configuration for tubes and micro-plates
13. Intuitive, user-friendly software interface
14. System should supplied with suitable PC
15. The system should be supplied with 3 years warranty. The system should be supplied with all the accessories required to function.
16. The system should be CE/ISO Certified.

