



Ref: INS/F-6110/2017-2018(N)

Date : 13/01/2018

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Biomedical refrigerator (As per specification attached)	Any	6.00	Nos.

Remarks: The Tenders can also to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/trade by RBI.

Note:inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose The offer should be submitted after fully considering the above notification.****

Please also indicate specifically the following :

1. The bids / quotation addressed to "**THE HEAD-PURCHASE**" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "**THE HEAD-PURCHASE**".
2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
3. Country of origion
4. Place of Shipment
5. Name, address,telephone number and fax numbers of your bankers.
6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
- 7.**Please Indicate frieght and insurance charges seperately.**
- 8.Please Indicate the Warranty Period and Delivery Period.
- 9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **01/02/2018 till 5.30**

Important Note

- 10.Shipping by AIRFREIGHT ONLY-CARGO MODE
- 11.Port of Entry for Customs clearance is -BANGALORE
- 12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: **INS/F-6110/2017-2018(N)**

Date : 13/01/2018

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

Namratha A

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



INSTITUTE FOR STEM CELL BIOLOGY AND REGENERATIVE MEDICINE

Ref No: INS/F-6110/2017-2018(N)

Specification of Biomedical Refrigerator

- 1 System should be vertical standing and should have working volume of 650 Ltrs±10%.
- 2 The system should be a research grade and suitable for continuous operation.
- 3 The temperature range of refrigerator should be 2°C to 23°C
- 4 The system should have LED/LCD display for temperature.
- 5 The system should have audio & visual alarms for low/ high temperature, door open and power failure. Preferably there should be an option to mute the audible alarms.
- 6 The system should have manual defrost.
- 7 The system should have single swing door.
- 8 The system should be high power efficient and have high grade thermal Insulation.
- 9 The system should have adjustable shelves.
- 10 The system should have Caster moving wheels with stopper for easy movement.
- 11 The system should have efficiency and low power consumption
- 12 Power supply should be 230V, 50HZ.
- 13 Power consumption for 24 hrs at normal room condition must be mentioned
- 14 The system should be supplied with stabilizer and all other accessories required for proper functioning of refrigerator.
- 15 The system should be certified for research grade as per norms of IS & CE or equivalent.
- 16 The Company must quote the freezer with all the suitable documents to justify the specification of product.
- 17 Must be offered with minimum 3-year warranty.
- 18 System should be compatible and have small foot print.
- 19 The company should have the service support in Bangalore

