



Ref: INS/F-6231/2017-2018(Y)

Date : 12/02/2018

ENQUIRY

Dear Sirs,

Please let us have your lowest Quote/Proforma Invoice indicating cost including packing ,insurance,estimated airfreight charges,handling charges,etc.CIF/CIP upto inStem Stores for the following materials.

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Liquid Nitrogen Container -60Ltrs Capacity (Specifications attached)	Any	1.00	Nos.

Remarks:The Tenders can also to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

NOTE:inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification.****

Please also indicate specifically the following :

1. The bids / quotation addressed to "THE HEAD-PURCHASE" with Tender / Enquiry Ref No and Item Description should reach us before submission Date and Time. The bids / quotation are liable to be rejected , if not addressed to "THE HEAD-PURCHASE".
2. **EXIM CODE -8 digit customs tariff Number -HSN (Harmonised System Nomenclature)**
3. Country of origion
4. Place of Shipment
5. Name, address,telephone number and fax numbers of your bankers.
6. Your Account number and swift number with the bank(required for payment purposes if the order materialises on your firm).
- 7.**Please Indicate frieght and insurance charges seperately.**
- 8.Please Indicate the Warranty Period and Delivery Period.
- 9.Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by **05/03/2018 till 5.30pm**

Important Note

- 10.Shipping by AIRFREIGHT ONLY-CARGO MODE
- 11.Port of Entry for Customs clearance is -BANGALORE
- 12.Please use International Freights which arrive directly into Bangalore.Ex.Lufthansa Airlines Cargo,Singapore Airlines Cargo.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/F-6231/2017-2018(Y)

Date : 12/02/2018

13.NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX,TNT, UPS,DHL ETC.SHIPMENTS SENT THROUGH FEDEX,TNT,UPS,DHL,ETC.WILL NOT BE ACCEPTED BY NCBS.

14.inStem is exempted in paying customs duty(except advolerum duty of 5% + 2% cess and 4% CVD on CIF value) PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK(CASH AGAINST DOCUMENTS).DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION.BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT.ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in blue ink, appearing to read 'Yesu R', written over a circular stamp.

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in



INSTITUTE FOR STEM CELL BIOLOGY AND REGENERATIVE MEDICINE

Ref No: INS/F-6231/2017-2018(Y)

Specification for 60L pressurized liquid nitrogen dewar

1. System should have all welded stainless steel construction compatible for LN2 storage
2. System should have a capacity of 60L with 5% error
3. System should be equipped with heavy duty castors with handles for easy mobility
4. System should have Twin liquid fill and decant ports
5. System should have an integral pressure build vaporizer with automatic pressure regulation
6. System should be equipped with electronic level gauge to give accurate reading of its contents
7. System should be equipped with a safety dispensing head incorporating fill, vent and dispense valves, safety relief valves and pressure gauges
8. Evaporation rate of the system should be less than 0.8%
9. The maximum allowable working pressure should be 3BarG
10. System should have ½” BSPPF outlet
11. System should include all other accessories including flexible hose pipes for dispensing liquid nitrogen
12. All the valves should be color coded for easy identification
13. System should have 5 years vacuum warranty
14. System should have 3 years of warranty
15. System should have complied with pressure equipment directive or equivalent
16. System should be ISO/CE certified
17. Installation of the LN2 dewar should be part of the offer

