Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5513/2017-2018(Y)

Date: 22/09/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Ouotation for the following:

SI.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Gel Documentation system (As per specification attached)	7	1.00	No.

Remarks:

Note: Quotation in USD, GBP, YEN, EURO, CND, SGD, ect will be acceptable.

1.	The	bids	shall	be	enclosed	in	an	envelope	,	and	due	date	sealed	duly	marked	"Tender	for
				_" R	ef No :							Th	e bids sl	nould	be addres	sed and t	o be
ma	ailed to	" TH	E HE	AD-I	PURCHA	SE"	. Th	e bids are l	iab	ole to l	be reje	ected i	f the sea	aled en	velope is	not addre	ssed
to	"THE	HEA	D-PU	RCF	IASE" w	ith T	ende	er Ref No	anc	l Item	Desc	riptio	n and di	ie date	. The bid	s delivere	ed in
pe	rson s	hall b	e dropj	ped i	n Purchas	se Se	ectio	n. If the b	ids	are s	ent th	rough	courier	or ma	ail , it sho	ould reacl	h by
su	bmissi	on Da	te and	Time	and inSte	em w	zill n	ot be respo	ns	ible fo	or the	delay.					-

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 11/10/2017 till 5.30 p.m.

- 3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.
- 4. The Validity of your quotation should be for 60 days from the date.
- 5.All duties,taxes,surcharge and cess as currently applicable must be sated in your quotation,seperately. Otherwise your quote is liable to be rejected.
- 6. Your quotation should indicate delivery period & Warranty period.
- 7.Delivery to be made to our Stores.Please indicate charges,if any extra.Transit Insurance should be done upto inStem Stores.
- 8.If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, Please state so and then offer alternative to quality/Specifications.
- 9. Payment: within one month after delivery & acceptance/satisfactory installation.
- 10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.
- 11.Since we are a public funded research institution,we are exempted from paying Customs Duty(Except advolerum duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments,technical instruments,equipments(including computers),their accessories,spares,consumables and software.Hence,please offer your prices taking this option into consideration.



Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5513/2017-2018(Y)

22/09/2017 Date:

12.If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate . contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Concillation Act, 1996 as amended from time to time shall apply. The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

> Yours faithfully For and on behalf of Insitute For Stem Cell Biology and Regenerative Medicine



GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No.:

91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax: 91-80-23636662 Website:

www.ncbs.res.in





Institute For Stem Cell Biology And Regenerative Medicine

Ref No: INS/L-5513/2017-2018(Y)

Specification for Gel-documentation system

The system should be suitable for analyzing gel with chemiluminescence, fluorescence, and digitization mode.

The system should have different mode of operation (standard/programmed)

The system should have Peltier cooled CCD camera.

The system should have scientific grade camera with a resolution of minimum 7 mega pixels

The system should have 16-bit grey scales

The system should have Motorized zoom option.

The system should have UV and white light trans-illuminator

The system should have different Epi light

The system should have minimum 5 filter option like GFP, Sybr, Ethidium bromide filter etc.

The system should have option for automatic and manual exposure.

The system should have adjustable manual exposure time.

The system should have auto and manual focusing

The system should have brightness control.

The system should be supplied with all the accessories required for calibrating the instrument.

The system should have sample trays for UV, white, Epi light.

The system should have UV safety shut off door system.

The system should be supplied with the analysis software.

The system should have option for 3d imaging.

The system should have LCD touch display.

The system should be supplied with suitable PC required for image analysis.

The system should operate on 230V,50Hz.

The system should be CE/ISO certified.

The system should have warranty of 3 years.

