



Ref: INS/L-5265/2017-2018(Y)

Date : 28/07/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		Supply and installation of SEM probe Current meter (As per specification attached)	Any	1.00	No.

Remarks :

Note :Quotation in USD, GBP, YEN, EURO, CND, SGD, ect will be acceptable.

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for _____ " Ref No : _____. The bids should be addressed and to be mailed to "**THE HEAD-PURCHASE**". The bids are liable to be rejected if the sealed envelope is not addressed to "**THE HEAD-PURCHASE**" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS
07/08/2017 till 5.30 p.m.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5.All duties,taxes,surcharge and cess as currently applicable must be sated in your quotation,seperately.Otherwise your quote is liable to be rejected.

6.Your quotation should indicate delivery period & Warranty period.

7.Delivery to be made to our Stores.Please indicate charges,if any extra.Transit Insurance should be done upto inStem Stores.

8.If you are unable to supply the quality,specifications or brand as mentioned in our enquiry,Please state so and then offer alternative to quality/Specifications.

9.Payment :within one month after delivery & acceptance/satisfactory installation.

10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11.Since we are a public funded research institution,we are exempted from paying Customs Duty(Except advolerum duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments,technical instruments,equipments(including computers),their accessories,spares,consumables and software.Hence,please offer your prices taking this option into consideration.

Institute For Stem Cell Biology and Regenerative Medicine
Autonomous Institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5265/2017-2018(Y)

Date : 28/07/2017

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'Yesu R', is written over a horizontal line.

Yesu R


GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666443/344344/345

Fax: 91 80 23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



Ref No: INS/L-5265/2017-2018(Y)

Specification for SEM Probe Current Meter

- 1) Meter should 3½-digit liquid crystal display readout of Electron Beam Probe Current.
- 2) Meter should free-standing and battery powered should measures current ranges from 1pA to 2μA in 5 switched ranges.
- 3) Meter should have automatic zero correction.
- 4) Meter input should be designed to accept large overloads in excess of 100μA without sustaining damage.
- 5) The Accuracy of the meters is
199.9pA accuracy +/- 5% full scale
1.999nA accuracy +/- 5% full scale
19.99nA accuracy +/- 2% full scale
199.9nA accuracy +/- 2% full scale.
1.999μA accuracy +/- 2% full scale.
- 6) Maximum acceptable error is 5pA at 25°C.
- 7) Power: Battery operated, life of the Battery should be 10,000 hours of operation
- 8) Meter should be supplied with all the accessories required for the operation.
- 9) Installation should be the part of the order.
- 10) System should have 1 year Warranty.
- 11) System should be ISO/CE certified.

