



Ref: INS/L-5746/2017-2018(Y)

Date : 14/11/2017

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		RO water system (Sprifications Attached)	any	1.00	No.

Remarks :

Note : The Tenders to be quoted in foreifn currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for _____ " Ref No : _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS
30/11/2017 till 5.30 p.m.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5.All duties,taxes,surcharge and cess as currently applicable must be sated in your quotation,seperately.Otherwise your quote is liable to be rejected.

6.Your quotation should indicate delivery period & Warranty period.

7.Delivery to be made to our Stores.Please indicate charges,if any extra.Transit Insurance should be done upto inStem Stores.

8.If you are unable to supply the quality,specifications or brand as mentioned in our enquiry,Please state so and then offer alternative to quality/Specifications.

9.Payment :within one month after delivery & acceptance/satisfactory installation.

10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11.Since we are a public funded research institution,we are exempted from paying Customs Duty(Except advolerum duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments,technical instruments,equipments(including computers),their accessories,spares,consumables and software.Hence,please offer your prices taking this option into consideration.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous Institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5746/2017-2018(Y)

Date : 14/11/2017

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'Yesu R'.

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA


Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in





inStem

Institute for Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Dept of Biotechnology, Govt. of India.

INS/F-5746/2017-2018(Y)

Reverse osmosis (RO) systems

1. Should be capable of delivering up to 100 L/h of purified water per hour
2. It should be an efficient, reliable, low-maintenance RO system that is affordable as well.
3. The high-performance RO membranes reduce water wastage and provide excellent product water quality, whilst ensuring high retention rates of water constituents.
4. Feed water for glassware machines, autoclaves and air humidifiers
5. Feed water for ultrapure water purificationsystems (ASTM type 1)
6. High performance, low-energy, thin-film composite (TFC) RO membranes
7. Feed pump which ensures stable pressure and water flow
8. 5- μ m particle filter
9. Integrated 100-L permeate tank
10. Compact design
11. High-quality integrated components
12. Fully assembled control panel.(No additional electrical installations required.)
13. CE-certified
14. Flow meter and manometer for control of permeate flow and pressure
15. Integrated pressure sensor with alarm function
16. Optional UV lamp and conductivity meter
17. Capacity : 100L
18. Electric connection : 230V
19. System should have 10 μ ms prefilters

