



Ref: INS/L-6122/2017-2018(N)

Date : 29/01/2018

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		4 ft chemical fume hood (specifications attached)	Any	1.00	No.

Remarks: The Tenders can also to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

NOTE: **inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification.**

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for _____" Ref No : _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS
15/02/2018 till 5.30pm**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5.All duties,taxes,surcharge and cess as currently applicable must be sated in your quotation,seperately.Otherwise your quote is liable to be rejected.

6.Your quotation should indicate delivery period & Warranty period.

7.Delivery to be made to our Stores.Please indicate charges,if any extra.Transit Insurance should be done upto inStem Stores.

8.If you are unable to supply the quality,specifications or brand as mentioned in our enquiry,Please state so and then offer alternative to quality/Specifications.

9.Payment :within one month after delivery & acceptance/satisfactory installation.

10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11.Since we are a public funded research institution,we are exempted from paying Customs Duty(Except advolerum duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments,technical instruments,equipments(including computers),their accessories,spares,consumables and software.Hence,please offer your prices taking this option into consideration.



Institute For Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-6122/2017-2018(N)

Date : 29/01/2018

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

Namratha A

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



INSTITUTE FOR STEM CELL BIOLOGY AND REGENERATIVE MEDICINE

Ref No: INS/L-6122/2017-2018(N)

Specifications for Fume Hood

1. The system should be a 4ft chemical fume hood and have the inner dimension of 1200 X 600 X 650 mm.
2. The system should be a powder coated steel exterior.
3. The working surface shall be molded epoxy or stainless steel.
4. The interior lining of the hood must be resistant to the materials and chemicals to which it will be exposed
5. The fume hood shall be equipped with a device to measure and monitor airflow.
6. The system shall have a visual indicator of the hood face velocity and adjustable low flow/caution alarm set points with audible buzzer or alarm
7. All interior surfaces of the hood shall be smooth and seamless, and constructed for easy cleaning.
8. Chemical storage drawer should be provided below the working bench with sufficient exhaust.
9. Hoods shall be equipped with sufficient fluorescent lighting. The light fixture should be easily accessible from the outside of the hood
10. The system should have digital display of air flow and alarms for low air suction and blockage on the exhaust outlet.
11. The system should have spill stopper work surface
12. The system should have front sliding toughened glass door.
13. The system should have water tap along with sink and drain port inside the cabinet.
14. System should have duplex electrical outlet points.
15. Plumbing fixtures shall be brass, chrome-plated, or acid and organic vapor resistant plastic. All fixtures shall have color-coded end caps
16. The system should have 4 gas line (2 gases, 2vacuum) ports.
17. The system should be supplied with all the accessories required to function.
18. The system should be quoted with exhaust blower with air suction capacity of 2500CFM.
19. The system should be supplied with the three phase starter
20. The system should have separate control switch to operate Blower motor.
21. The operating voltage of the system should be 230V, 50Hz
22. The system should be quoted with the FRP ducting.
23. The noise level of the system should be <70dB.
24. The system should have warranty of 3 years.
25. The system should be CE/ISO certified and should meet all the standard certification.

