

ENQUIRY

Dear Sirs,
Please let us have your lowest Quotation for the following :

Sl.No	Cat.No	Item Description	Make/Model	Item Qty	UOM
1		- 20 to-30 degree freezer (specifications attached)		1.00	No.

Note: 2 PART TENDER: The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.

1. The bids shall be enclosed in an envelope , and due date sealed duly marked "Tender for _____" Ref No : _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail , it should reach by submission Date and Time and inStem will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS
28/08/2019 till 5.30 p.m.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done upto inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, Please state so and then offer alternative to quality/Specifications.

9. Payment : within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except ad valorem duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.



Institute For Stem Cell Biology and Regenerative Medicine
Autonomous institute of the Department of Biotechnology, Government Of India



Ref: INS/L-5324/2019-2020(Y)

Date : 13/08/2019

12.If the item is covered under DGS&D rate contract,please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees.The decision of the arbitrator shall be final and binding on the parties.The venue for arbitration shall be Bangalore.The provisions of the Arbitration and Concillation Act,1996 as amended from time to time shall apply.The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Insitute For Stem Cell
Biology and Regenerative Medicine

A handwritten signature in black ink, appearing to read 'Yesu R', with a small 'R' as a superscript.

Yesu R

GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Fax : 91-80-23636662

Email Id: purchase@ncbs.res.in

Website : www.ncbs.res.in



2 PART TENDER FOR -20/-30 Freezer Technical Specifications -Qty 01 No.

P.O. No: INS/L-53242019-2020(Y)

Tender Specifications for -20/-30 Freezer

Should have a Temperature range: -20°C to - 30°C

Should have a cooling coils strategically spaced in the walls, not just along both sides and the back of the chamber, but also on the top and the bottom of the chamber to ensure maximum cooling performance

Maximum cooling performance: -30 °C (Ambient Temperature: 35°C)

Effective Capacity: 24.4 Cu ft (690L)

Large-Capacity biomedical freezer with direct cooling and manual defrost

External Dimensions (W x D x H) - 770 x 830 x 1955 mm

Internal Dimensions (W x D x H) - 650 x 700 x 1520 mm

Temperature Control: Microprocessor Control

Alarm and Safety: High/Low Temperature alarm, power failure, Door ajar alarm remote alarm contacts (DC 30V , 2A) , Memory Back up

Digital Display of temperature

Should have a Direct Cooling System

Should have a high freezing Power

Should have SS Chamber and Exterior with painted steel to avoid corrosion and starches

Insulation: HCFC - free rigid polyurethane foamed in place

Compressor: Hermetic rotary type (400W)

Refrigerant: HFC

Shelves: 4 adjustable wire shelve, 50 kg capacity each

Access Ports: Diameter 30 mm (left) / Diameter 20mm (top)

Temperature Range: -18°C to -30°C (Ambient Temperature: 35°C)

Casters: 4 Casters with 2 adjustable feet

Accessories: 1 Key set, 1 set scraper, 2 large & 2 large small nylon clips

Should be a imported instrument with all below certifications

Quality management system: ISO9001

Medical Device Management System: ISO13485

Environmental Management System: ISO14001

CE Certified

Warranty for 3 years



Terms and Conditions:

1. The Technical and Financial / Price Bids shall be submitted simultaneously in two (2) cover (sealed) system. The proposals shall be evaluated in two stages: (1) Technical and (2) Price / Financial. Technical evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial Proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.
2. **first sealed cover – Cover I**, and super scribed as **“Techno-commercial Bid”** and should contain Complete Technical details of the Instrument offered (Specifications, Technical Parameters, Advantages, etc.,)
3. The **second sealed cover – Cover II** super scribed **'Price Bid'** should contain **only rates** (should be duly signed with seal and filled with date wherever necessary)
4. THESE TWO COVERS SHALL BE AGAIN PUT INTO A SINGLE WAX SEALED COVER super scribed **“-20/-30 Freezer -Qty 01 No.”** and should reach **INSTEM on or before 28-08-2019 before 17.30 hrs”**. This should be addressed to the Purchase Officer, inStem, NCBS, GKVK Post, Bellary Road, Bangalore – 65.
5. The tender to be quoted in foreign currencies & any other currencies approved/traded by RBI- USD/Euro/JPY/GBP/SGD/CAD/INR.
6. If the items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.



INFORMATION TO TENDERERS

The Tender shall be evaluated under 2 (Two) Bid System

I Technical Bid

II Financial Bid

TECHNICAL SPECIFICATIONS & EVALUATION CRITERIA WITH MARKS FOR 2 PART TENDER FOR “-20/-30 Freezer -Qty 01 No”.

Sl No.	Tender Specifications	Marks
1	Should have a Temperature range : -20°C to - 30°C	5
2	Should have a cooling coils strategically spaced in the walls, not just along both sides and the back of the chamber, but also on the top and the bottom of the chamber to ensure maximum cooling performance	
3	Maximum cooling performance : -30 °C (Ambient Temperature: 35°C)	
4	Effective Capacity : 24.4 Cu ft (690L)	10
5	Large-Capacity biomedical freezer with direct cooling and manual defrost	
6	External Dimensions (W x D x H) - 770 x 830 x 1955 mm Internal Dimensions (W x D x H) - 650 x 700 x 1520 mm	
7	Temperature Control : Microprocessor Control	5
8	Alarm and Safety : High/Low Temperature alarm, power failure, Door ajar alarm remote alarm contacts (DC 30V , 2A) , Memory Back up	5
9	Digital Display of temperature	5
10	Should have a Direct Cooling System	5
11	Should have a high freezing Power	
12	Should have SS Chamber and Exterior with painted steel to avoid corrosion and starches	5
13	Insulation: HCFC - free rigid polyurethane foamed in place	5
14	Compressor : Hermetic rotary type (400W)	5
15	Refrigerant : HFC	5
16	Shelves : 4 adjustable wire shelve, 50 kg capacity each	10
17	Access Ports: Diameter 30 mm (left) / Diameter 20mm (top)	5
18	Temperature Range : -18°C to -30°C (Ambient Temperature : 35°C)	10
19	Casters : 4 Casters with 2 adjustable feet	5
20	Accessories : 1 Key set, 1 set scraper, 2 large & 2 large small nylon clips	5
21	Should be a imported instrument with all below certifications	5
22	Quality management system : ISO9001	
23	Medical Device Management System : ISO13485	
24	Environmental Management System : ISO14001	
25	CE Certified	5
26	Warranty for 3 years	
	Total	100

Evaluation will be carried out and those Vendors who score minimum 75% will qualify for Price Bid opening. Thereafter, Financial proposal shall be evaluated. The Commercially LOWEST BIDDER shall be the first preferred Vendor for award of Order.