

Ref: INS/L-5588/2018-2019(Y)

Date: 26/09/2018

ENQUIRY

Dear Sirs,

Please let us have your lowest Quotation for the following:

SL. No	Cat. No	Item Description	Make/Model	Item Qty	UOM
1		Water Jacketed co2 Incubator with cooling capacity (Specifications attached)		1.00	Nos.

Remarks: The Tenders can also to be quoted in foreign currencies- USD/Euro/JPY/GBP/SGD/CAD or any other currencies approved/traded by RBI.

NOTE: **inStem is a public funded research institute and is entitled to concessional rate of GST @ 5% for items supplied for research purpose. The offer should be submitted after fully considering the above notification. **

1. The bids shall be enclosed in an envelope, and due date sealed duly marked "Tender for _____" Ref No: _____. The bids should be addressed and to be mailed to "THE HEAD-PURCHASE". The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission Date and Time and inStem will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS

Till 03/10/2018 5.30 p.m.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The Validity of your quotation should be for 60 days from the date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & Warranty period.

7. Delivery to be made to our Stores. Please indicate charges, if any extra. Transit Insurance should be done up to inStem Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/Specifications.

9. Payment :within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is super scribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs Duty (Except advalorem duty of 5% + 2% cess and CVD @4% vide Notification No.51/96 with latest amendments) and excise duty vide Notification No.10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipment, technical instruments, equipment (including computers), their accessories, spares, consumables and software. Hence, please offer your prices taking this option into consideration.



Institute for Stem Cell Biology and Regenerative Medicine

Autonomous institute of the Department of Biotechnology, Government of India



Ref: INS/L-5588/2018-2019(Y)

Date: 26/09/2018

12.If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13.Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act,1996 as amended from time to time shall apply. The Courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

Yours faithfully

For and on behalf of Institute for Stem Cell
Biology and Regenerative Medicine


Yesu R



GKVK, Bellary Road, Bangalore-560065,INDIA

Phone No. : 91-80-23666343/344/345/346

Email Id: purchase@ncbs.res.in

Fax : 91-80-23636662

Website : www.ncbs.res.in



Ref No: - INS/L-5588/2018-2019(Y)

Specifications for Water-jacketed CO2 Incubator with cooling capacity.

1. The system should be Water jacketed Co2 incubator.
2. The system should be microprocessor controlled Co2 incubator. It should have touch screen LCD controller with easy icons and user-friendly control system.
3. The system should have the capacity of 190 L ($\pm 5\%$)
4. The system should be equipped with large, easy to read display screen.
5. The system should be supplied with 3nos of adjustable SS shelves
6. The system should have two doors and inner door should be glass door to seal the chamber with Silicon/Rubber gasket.
7. The fluctuation/variation at 37°C should not be more than 0.1°C
8. The system should have the operating temperature range from -10°C to 60°C. The temperature accuracy and uniformity should be less than 0.1°C.
9. The system should have a cooling coil which should be factory installed with the chiller unit.
10. The system should have the Infrared CO2 sensor.
11. The system should have the CO2 range 1-20% with accuracy 0.1%.
12. The system should have the option to mute the audible alarms.
13. The system should have the seamless SS inner chamber.
14. The system should have onboard Diagnostic interface to show system parameters and functions.
15. The system should be offered with power rating of single phase 230 V 50 Hz.
16. The system should be equipped with buzzer and visual alarm for power failure, CO2, and temperature deviation against the setup temperature.
17. The system should be supplied with all the accessories required to function.
18. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).
19. The system should be supplied with 3-year warranty as a part of the offer.
20. The System should be CE/ISO Approved.
21. The company should have base in Bangalore

